



KERALA GAZETTE

കേരള ഗസറ്റ്

PUBLISHED BY AUTHORITY

ആധികാരികമായി പ്രസിദ്ധപ്പെടുത്തുന്നത്

Vol. II വാല്യം 2	THIRUVANANTHAPURAM, TUESDAY തിരുവനന്തപുരം, ചൊവ്വ	31st December 2013 2013 ഡിസംബർ 31 10th Pousha 1935 1935 പൗഷം 10	No. നമ്പർ	52
---------------------	---	--	--------------	----

PART III

Stores Purchase

Insurance Medical Services Department

TENDER NOTICE

(e-Government Procurement)

No. F4/13778/2013/DIMS. 4th December 2013.

The Director of Insurance Medical Services, Kerala, invites online bids from reputed manufactures/authorised dealers for the supply & installation of CT Scanner (1 No.) at ESI Hospital, Feroke, Kozhikode.

Tender No.—01/2013-14.

Estimated Cost—₹ 1,00,00,000.

EMD—₹ 1,00,000.

Tender Fee—₹ 15,750.

Instructions to bidders

1. *Mode of submission of bids:*—All bid documents shall be submitted online through the web site www.etenders.kerala.gov.in in their designated online covers. No other mode of submission shall be accepted and such tenders will be rejected outright.
2. *Cover details:*—The tenders should be submitted online in two separate parts.

(i) Technical Bid (Cover-I)

Contents:

- (1) Organization/Company profile.
- (2) Registration certificate
- (3) Detailed specification of the equipment/machine.
- (4) ISO/ISI mark certificate/ document from ICRD Country of origin
- (5) Radiation safety certificate
- (6) TIN certificate
- (7) Name and address of institutions in Kerala where the equipment have already been installed
- (8) PAN/TAN certificate as per requirement.

(ii) Financial Bid (Cover-II)

Contents:

- (1) Price bid (BOQ)

The price bid should be prepared by downloading the BOQ template published along with this tender, fillup the relevant information, digitally sign and upload

the same as per the Government of Kerala eTendering process. Bidder should strictly follow the technical formalities involved as prescribed by the Government of Kerala eTendering process. The price quoted should be inclusive of all taxes and duties.

- (1) AMC rates for 5 years after expiry of warranty- Comprehensive
- (2) AMC rates for 5 years after expiry of warranty—Non comprehensive

(2) *Downloading of e-Tender documents:*—The tender documents can be downloaded from the website www.etenders.kerala.gov.in from 5-12-2013, 3 p.m. onwards to 4-1-2014, 3 p.m.

(3) *Submission of e-Tender documents:*—The digitally signed tender documents shall be submitted online through the web site www.etenders.kerala.gov.in from 5-12-2013 to 5-1-2014, 3 p.m.

(4) *Payment of Tender Fees:*—A non-refundable tender fee of ₹ 15,750 inclusive of VAT. Payment should be made through online payment incorporated with the Kerala Government eTendering process (SBT or NEFT online payment system). No other mode of payment shall be accepted.

(5) *Payment of Earnest Money Deposit (EMD):*—An amount of ₹ 1,00,000 as EMD shall be paid online (SBT or NEFT online payment system) through the web site www.e-tenders.kerala.gov.in. No other mode of remittance shall be accepted.

(6) *Exemption from payment of EMD :*—Bidders who are exempted from furnishing Earnest money deposit have to upload the concerned document proof.

(7) *Opening of e-Tenders:*—The Technical and Financial bids will be opened by the Technical Committee and the Financial Committee in two sessions as per the tender Schedule.

(8) *Evaluation*

The Technical Evaluation Committee will scrutinize the Technical Bid in line with the Departmental specification. The technical bid which is cleared by Technical Evaluation Committee alone will be available to the Financial bid evaluation.

Technical Evaluation is followed by Financial Evaluation of the Financial evaluation committee, which in result, decides the selection of the bidder for the award of contract (AOC).

(9) *Award of Contract (AOC) :*— Successful bidder will be issued with the award of contract.

(10) *Note to Bidders:*—

- (i) Bidders have to procure legally valid Digital Certificate as per Information Technology Act, 2000 for digitally signing their electronic bids. For details regarding digital certificates, please visit the e-GP website www.etenders.kerala.gov.in.

- (ii) Bidders are advised to note the Tender-Id and Tender No. & Date for future reference.
- (iii) All uploaded documents should contain the signature and the office seal of the bidder/ authorized persons and should be digitally signed while uploading. Documents uploaded without digitally signing shall entitle rejection of the tender.
- (iv) For help on e-tendering process, contact Kerala State IT Mission, e-Government Procurement PMU & Helpdesk, Basement floor of Pension Treasury Building, Uppalam Road, Statue, Thiruvananthapuram; Ph: 0471-2577088, 2577188; e-mail: etendershelp@kerala.gov.in; Website: www.etenders.kerala.gov.in on all Government working days from 9.30 am to 5.30 p.m.
- (v) For any clarifications regarding the tender documents, specifications, terms and conditions and other matters, please contact The Office of the Directorate of Insurance Medical Services, Thycaud, Thiruvananthapuram-14.

Note:—Bidders are advised to go through the conditions of the tender notice carefully and comply them to avoid outright rejection of their tender.

DETAILED MACHINE SPECIFICATION

<i>Sl.No.</i>	<i>Specification</i>
(1)	(2)
1	<p>C. T. Scanner</p> <p>The model offered should be High-end model under current production, should be with Slip ring Technology. The offer should meet the specifications as follows:</p>
2	<p>X-Ray Section</p> <p>The X-Ray generator should be compact and inbuilt in the Gantry.</p> <p>The mA range available to be up to 200 mA or more with increments in steps for not more than 10 mA</p> <p>The X-Ray tube should be essentially Dual focus with capacity of at least 2.0 MHU.</p>

(1)	(2)
	The X-Ray tube should have a cooling rate of not less than 700 Kwh per minute
	The X-Ray tube cooler should be built in the Gantry.
3.	Detector <p>The detector offered should be solid state/ceramic</p> <p>The effective elements/ channels should be at least 650 per row. It should have 2 or more slices per rotation in all slice thickness.</p> <p>Specify the Fan angle of the X-rays and the types of geometry</p> <p>The detectors should not require frequent calibration.</p>
4.	Gantry <p>The gantry should be provided with user - friendly control panels for easy positioning</p> <p>Multi slice facility of 2 or more slices should be standard for all slice thickness.</p> <p>Aperture should be at least 65 cms diameter</p> <p>The CT Scanner should have low voltage slip rings incorporated in the Gantry</p> <p>The scan time for a 360 Degree rotation should be a second or less.</p> <p>The gantry should have a minimum Physical tilt of 25 degrees on either side and remote tilt should be available as standard.</p>
5.	Patient Couch <p>The range of metal free scan should be at least 120 cms.</p> <p>The vertical range should be at least 5 cms (maximum height - minimum height)</p> <p>Specify the reproducing accuracy of the table</p> <p>Remote UP/DOWN, FORWARD, BACKWARD should be standard.</p> <p>The patient table offered should have a minimum load bearing capacity of at least 200 kg.</p>
6.	Spiral Section <p>The range of spiral facilities in Axial direction should be more than 100 cms</p> <p>The reconstruction time in spiral scan should not be more than 1.0 seconds</p>

(1)	(2)
7.	Computer Section <p>The computer offered should be the latest Multi tasking with 64 Bit processor and a menu driven platform.</p> <p>The monitor should be the latest colour of at least 18 inches and full flat screen</p> <p>The display matrix should be at least 1024 x 1024. The Hard Disk capacity for both image and raw data should be 240 GB or more</p> <p>It should have facility to store at least 10000 images.</p> <p>The system should be supported with archiving facility. CD and DVD archiving facility are required.</p> <p>DICOM facility to send, store, print, receive etc. should be standard.</p> <p>Fully DICOM 3.0 (or newer version if available at time of delivery)</p>
8.	Image Processing Section <p>The following software should be offered as standard (MPR, ROI, Volume Calculation, CT Number Display, Window Width, Window Level, To Program Display, Cine Display, HRCT/LUNG, CT Angiography)</p>
9.	Resolution <p>The system should have a resolution of at least 15 LP/cm at standard Parameters and using CAT Phantom</p>
10.	Dry imager/Camera <p>A dry imager with Digital Interface and Control integrated with man console, Dicom 3 (or newer version if available) compatible</p> <p>Camera should print on 14" x 17" film size</p> <p>Camera having dpi resolution of 500 dpi or more</p>
11.	Accessories <p>Lead glass of at least 2 ft by 4 ft</p> <p>Voltage Stabilizer of Suitable Capacity with 30 mt backup</p> <p>2 operator chairs</p> <p>Light weight Viny Lead Aprons of 0.5mm lead equivalence with thyroid guard AERB Approved - 2 no's</p>

Turnkey Work For Model

1. Minor Civil works to remodel the existing room.
2. Fixing of the lead glass on the wall.
3. Tile/Vinyl flooring on the gantry and console room.

- * Interior finishing works with false ceiling.
- * Dadoing work till false ceiling in the console and gantry room.
- * Supply and Installation of the AC units of suitable tonnage on the gantry and console room.
- * Necessary cabling work, lighting, switches and wiring sockets and electrical distribution paneling works inside the C. T. room.
- * Lead lining on the doors and frames as per AERB requirements.
- * Obtaining DRS approved by liaisoning with hospital authorities (by supplier).
- * Obtaining electrical inspectorates approval (by supplier) by liaisoning with hospital authorities.
- * The cost of turnkey work should be specified separately as optional.

TERMS AND CONDITIONS

1. All conditions mentioned in the Kerala Store Purchase Manual are applicable to this tender.
2. Tenderers have to follow the e-Payment mechanism of e Tendering process for the payment of EMD and Tender fee.
3. Delivery shall be made within 30 days from the date of award of contract.
4. The item should be delivered/installed at the ESI Hospital, Feroke, Kozhikode as the case may be, at the risk and cost of the supplier.
5. The equipment must be a newly manufactured item and not a reconditioned one.
6. The item shall carry a clear warranty for a period of 3 years from the date of delivery/installation and commissioning as the case may be.
7. The supplier shall execute Annual Maintenance Contract for a period of 5 years on expiry of the warranty period. Hence, both comprehensive and non comprehensive AMC rates for 5 years after the warranty period for each item has to be quoted separately.
8. The manufacturer/supplier shall have their own service facilities in Kerala.

9. The item shall carry ISI mark/ISO mark certificate/ Radiation Safety Certificate from ICRD/Country of origin as the case may be, wherever applicable. Attested copies of these certificates issued by the Competent Authority are also to be submitted with the tender.
10. Name and address of a few Institutions in Kerala where the equipments have already been supplied/ installed by the bidder shall be furnished along with the Tender.
11. As and when necessary, one employee of this Department shall be given training in use of this equipment by the supplier at his cost.
12. Payment shall be made after installation and successful functioning for a continuous period of one month from the start date of working.
13. The bidder shall furnish attested copy of their registration certificate, TIN and TAN certificate etc issued by the competent authority.
14. The successful bidder shall deposit a sum equivalent to 5% of the value of the contract as Security Deposit, which will be released only after the guarantee period.
15. Hard copies of submitted Technical cover documents and other relevant documents should reach before the Director of Insurance Medical Services before bid opening date.
16. The decision of the Director of Insurance Medical Services will be final and reserves the right to accept or reject the tender without assigning any reasons there of as provided under Store Purchase Rules.

Directorate of Insurance
Medical Services, Thycaud,
Thiruvananthapuram.

(Sd.)
*Director of Insurance
Medical Services.*

ദർശനം പരസ്യം

നമ്പർ-എഫ്4-23578/2013/ഡി. ഐ. എം. എസ്.

2013 ഡിസംബർ 5.

ഇൻഷുറൻസ് മെഡിക്കൽ സർവീസസ് വകുപ്പിന്റെ കീഴിൽ പ്രവർത്തിക്കുന്ന 9 ഇ. എസ്. ഐ. ആശുപത്രികളിൽ 1-4-2014 മുതൽ 31-3-2015 വരെയുള്ള കാലയളവിലേക്ക് ഭക്ഷ്യ-ഭക്ഷ്യതര വസ്തുക്കൾ വിതരണം ചെയ്യുന്നതിന് മുദ്രവച്ച മൽസരാടിസ്ഥാനത്തിലുള്ള ദർശനങ്ങൾ ക്ഷണിച്ചുകൊള്ളുന്നു.

ദർഘാസുകൾ 2014 ജനുവരി 2-ാം തീയതി 2 മണിവരെ ബന്ധപ്പെട്ട ആശുപത്രി സുപ്രണ്ടുമാർ സ്വീകരിക്കുന്നതായിരിക്കും. അന്നേ ദിവസം ഉച്ചയ്ക്കുശേഷം പകൽ 3 മണിക്ക് അതേ ആഫീസിൽ ദർഘാസുകാരുടെയോ അവരുടെ അംഗീകൃത പ്രതിനിധിയുടെയോ സാന്നിധ്യത്തിൽ ദർഘാസുകൾ തുറക്കുന്നതായിരിക്കും.

ദർഘാസുകളുടെ വില ബന്ധപ്പെട്ട ആശുപത്രി മേധാവി സ്റ്റോർ പർച്ചേയ്സ് റൂൾ അനുസരിച്ച് നിശ്ചയിക്കുന്നതായിരിക്കും. “0210-01-101-99” റസീപ്റ്റ് ഫ്രം ഇ. എസ്. ഐ. സ്കീം എന്ന ഹെഡിൽ ടെണ്ടർ ഫോറത്തിന്റെ വിലയും “0040-00-102-99” എന്ന ഹെഡിൽ വിലപന നികുതിയും ട്രഷറിയിൽ അടച്ച രസീത് 2013 നവംബർ 1-ാം തീയതിക്കുശേഷമുള്ള ബന്ധപ്പെട്ട റവന്യൂ അധികാരിയിൽ നിന്നും ലഭിച്ച സോൾവെൻസി സർട്ടിഫിക്കറ്റിന്റെ സാക്ഷ്യപ്പെടുത്തിയ ശരിപകർപ്പ് ഒറിജിനൽ സഹിതം സുപ്രണ്ടിന് അപേക്ഷ നൽകി ആവശ്യമായ ദർഘാസ് ഫോറവും മറ്റ് അനുബന്ധ രേഖകളും വാങ്ങാവുന്നതാണ്. ദർഘാസ് ഫോറത്തിന്റെ വിൽപന 2013 ഡിസംബർ 19-ാം തീയതി തിങ്കളാഴ്ച തുടങ്ങുന്നതും 2013 ഡിസംബർ 31-ാം തീയതി ചൊവ്വാഴ്ച പകൽ 3 മണിക്ക് അവസാനിക്കുന്നതുമാണ്. പുരിപ്പിച്ച ദർഘാസുകൾക്കൊപ്പം കോൺട്രാക്ടർ/ദർഘാസുകാർ അവർക്ക് ബന്ധപ്പെട്ട അധികാരികളിൽ നിന്നും ലഭിച്ച ഫുഡ് ലൈസൻസിന്റെ സാക്ഷ്യപ്പെടുത്തിയ ശരിപകർപ്പ് ഉൾപ്പെടെയുള്ള മറ്റ് അനുബന്ധ രേഖകളും സമർപ്പിക്കേണ്ടതാണ്. ഭക്ഷ്യ സാധനങ്ങൾക്കും ഭക്ഷ്യതര സാധനങ്ങൾക്കും പ്രത്യേകം ദർഘാസുകൾ വേണ്ടതാകുന്നു. കൂടുതൽ വിവരങ്ങൾ ബന്ധപ്പെട്ട സ്ഥാപനങ്ങളിൽ നിന്നും പ്രവർത്തി സമയങ്ങളിൽ ലഭിക്കുന്നതാണ്.

ദർഘാസുകൾ സ്വീകരിക്കുന്നതിനോ നിരാകരിക്കുന്നതിനോ ഉള്ള അധികാരം ബന്ധപ്പെട്ട അധികാരികളിൽ നിക്ഷിപ്തമാണ്.

ആശുപത്രികളുടെ വിവരം

1. ഇ. എസ്. ഐ. ആശുപത്രി, പേരൂർക്കട, തിരുവനന്തപുരം.
2. ഇ. എസ്. ഐ. ആശുപത്രി, വടവാതൂർ, കോട്ടയം.
3. ഇ. എസ്. ഐ. ആശുപത്രി, ആലപ്പുഴ.
4. ഇ. എസ്. ഐ. ആശുപത്രി, എറണാകുളം
5. ഇ. എസ്. ഐ. ആശുപത്രി, ഒളിക്കര, തൃശ്ശൂർ.
6. ഇ. എസ്. ഐ. ആശുപത്രി, മുളംകുന്നത്തുകാവ്, തൃശ്ശൂർ.
7. ഇ. എസ്. ഐ. ആശുപത്രി, പാലക്കാട്.
8. ഇ. എസ്. ഐ. ആശുപത്രി, ഫറോക്ക്, കോഴിക്കോട്.
9. ഇ. എസ്. ഐ. ആശുപത്രി, തോട്ടട, കണ്ണൂർ.

ഡയറക്ടറേറ്റ് ഓഫ്	(ഒപ്പ്)
ഇൻഷുറൻസ് മെഡിക്കൽ	ഇൻഷുറൻസ് മെഡിക്കൽ
സർവ്വീസസ്, തിരുവനന്തപുരം.	സർവ്വീസസ് ഡയറക്ടർ.